

TOTAL PROTEINS (BIURET)

Colorimetric determination of proteins in serum and plasma

TEST SUMMARY

Proteins react with copper ions forming a red compound measurable spectrophotometrically.

SAMPLES

Serum or plasma.

Stability: 1 month at 2-8°C.

REAGENTS

Standard:

Sole Reagent: Copper sulphate mM; Potassioum ioduro 30mM; Sodium potasium tartrato 18 mM;

Sodium hydroxide 20 mM.

Proteins 5 g/dl; stabilizers and preservatives.

MATERIAL REQUIRED BUT NOT SUPPLIED

Normal laboratory equipment. Spectrophotometer thermostatation. Automatic with Micropipette. Cuvette in optical glass or monouse in optical polystyrene. Physiologic solution.

PRECAUTIONS

Reagent may contain not reactive conservative components. It is opportune to avoid contacts with the skin and do not swallow

Perform the test according to the general "Good Laboratory Practice" (GLP) guidelines.

REAGENTS PREPARATION

Reagents are ready to use, and are stable until expiration date on label and are stable until expiration date on label

Sole Reagent is stored at room temperature, while the standard at 2-8°C

Pay attention, do not contaminate reagents when vials are open.

PROCEDURE

End Point Kind of analysis: Reading time: 5 minutes Colour stability: 60 minutes 546 nm (520-570) Wavelength:

Temperature: 37°C Lightpath: 1 cm

Zero: Blank Reagent

REAGENTS	BLANK	STANDARD	SAMPLE
Sole Reagent	1 ml	1 ml	1 ml
Distilled Water	10 μl		
Standard		10 μl	
Sample			10 μl

Mix and incubate at 37°C at least for 5 minutes. Read the absorbance against Blank at 546 nm. Colour is stable 60 minutes

CALCULATION

A (sample) Proteins g/dl x 5 A (standard)

EXPECTED VALUES

6.3 - 8.2 g/dl

Every laboratory should establish own reference intervals in accordance with own population.

NOTES

- If the results are incompatible with clinical presentation, they have to be evaluated within a total clinical study.
- Only for IVD use.

CALIBRATION/ QUALITY CONTROL

It is suggested to perform an internal quality control. For this purpose the following control sera on human base are available on request:

10 x 5 ml

QN 0050 CH 10 x 5 ml Control Sera normal values

Control Sera pathological values

TEST PERFORMANCE

Precision

QP 0050 CH

Intra-assay (n = 20)	Mean (g/dl)	SD (g/dl)	CV%
Sample 1	6.2	0.032	0.52
Sample 2	6.8	0.032	0.48

Inter-assay (n = 20)	Mean (g/dl)	SD (g/dl)	CV%
Sample 1	6.195	0.039	0.64
Sample 2	6.81	0.044	0.66

Methods comparison

A comparison with a commercial available product gave the following results in a comparison on 50 samples:

Total proteins LTA = x Total proteins competitor = y

y = 0.96976x + 0.20025r = 0.99128

WASTE DISPOSAL

Product is intended for professional laboratories. Waste products must be handled as per relevant security cards and local regulations.

PACKAGING

CODE CC02100 (600 TESTS) Sole Reagent 6 x 100 ml (liquid) 1 x 3 ml Standard (liquid)

REFERENCES

Dumas B.T. e coll. - Clin. Chem. 27,; 16472 (1981)

MANUFACTURER

LTA s.r.l. Via Milano 15/F

20060 Bussero (Milan) ITALY ++39 02 95409034 Fax: ++39 02 95334185 e-mail: info@ltaonline.it Website: http://www.ltaonline.it

SYMBOLS

IVD Only for IVD use

LOT Lot of manufacturing

REF Code number

X Storage temperature interval

Expiration date

⚠ Warning, read enclosed documents

 \prod i Read the directions Biological risk

Mod. 01.06 (ver. 4.2 - 04/03/2006)

