



# GALACTOSE – CONTROL SET

## Control solutions for determination of Galactose in urine

### INTENDED USE

Concentrations and activities of components are studied for an excellent use even in automatic system of clinical chemistry, even in manual procedures.

### PRESERVATION AND STABILITY

Preserve the product at 2-8°C, stable until expiration date on the package.  
Once opened controls are stable for 307 days at 2-8°C.

### MATERIALS REQUIRED BUT NOT SUPPLIED

Normal laboratory equipment.

### REAGENTS PREPARATIONS

Controls are ready to use.  
Bring controls at room temperature before use.

### PRECAUTIONS

- Product contains Sodium Azide 0.095% as preservative.
- Avoid contact with skin and mucous membrane.
- Use the normal precautions for lab behaviour.
- For in vitro use only.

### WASTE DISPOSAL

Product is intended for professional laboratories. Waste products must be handled as per relevant security cards and local regulations.

### PACKAGING

#### CODE CC03830

Control level 1            2 x 1 ml    (liquid)  
Control level 2            2 x 1 ml    (liquid)

### REFERENCES

Archives data.

### CALIBRATION VALUES

E317.9                      2013 - 11

COMPONENT / METHOD	VALUE	RANGE	UNIT
Enzymatic test for determination of Galactose in urine			
CONTROL LEVEL 1 C317.9 2013 - 11	10.24	8.39 – 12.08	mg/dl
CONTROL LEVEL 2 D317.9 2013 - 11	50.02	41.01 – 59.02	mg/dl

### MANUFACTURER

LTA s.r.l.  
Via Milano 15/F  
20060 Bussero (Milan) ITALY  
Tel: ++39 02 95409034  
Fax: ++39 02 95334185  
e-mail: info@ltaonline.it  
Website: <http://www.ltaonline.it>

### SYMBOLS

- Only for IVD use
- Lot of manufacturing
- Code number
- Storage temperature interval
- Expiration date (year – month)
- Warning, read enclosed documents
- Read the directions
- Biological risk

Mod. 01.06 (ver. 1.1 - 17/07/2006)

