

IODIDE – CONTROL SET

Control solutions for determination of Iodide in urine

INTENDED USE

Concentrations and activities of components are studied for an excellent use even in automatic system of clinical chemistry, even in manual procedures.

PRESERVATION AND STABILITY

Preserve the product at 2-8°C, stable until expiration date on the package.
Once opened controls are stable for 7 days at 2-8°C.

MATERIALS REQUIRED BUT NOT SUPPLIED

Normal laboratory equipment.

REAGENTS PREPARATIONS

Controls are ready to use.
Bring controls at room temperature before use.

PRECAUTIONS

- Product contains Sodium Azide 0.095% as preservative.
- Avoid contact with skin and mucous membrane.
- Use the normal precautions for lab behaviour.
- For in vitro use only.

WASTE DISPOSAL

Product is intended for professional laboratories. Waste products must be handled as per relevant security cards and local regulations.

PACKAGING

CODE CC02830





Control level 1 1 x 2 ml (liquid)
Control level 2 1 x 2 ml (liquid)

REFERENCES

Archives data.

CALIBRATION VALUES









 D054.18  2020/02

COMPONENT / METHOD	VALUE	RANGE	UNIT
Enzymatic test for determination of Galactose in urine			
CONTROL LEVEL 1			
 A044.18  2020/02	38.7	31.7 – 45.6	µg/l
CONTROL LEVEL 2			
 B044.18  2020/02	226.9	186.1 – 266.8	µg/l

MANUFACTURER

LTA s.r.l.
Via Milano 15/F
20060 Bussero (Milan) ITALY
Tel: +39 02 95409034
Fax: +39 02 95334185
e-mail: info@ltaonline.it
Website: www.ltaonline.it

SYMBOLS

-  Only for IVD use
-  Lot of manufacturing
-  Code number
-  Storage temperature interval
-  Expiration date (year – month)
-  Warning, read enclosed documents
-  Read the directions
-  Biological risk

Mod. 01.06 (ver. 1.0 - 17/04/2012)

